



Mercodia Animal Insulin Control (Low, Medium, High)

Directions for Use

10-1221-01

For Research Use Only

Manufactured by

Mercodia AB
Sylveniusgatan 8A
SE-754 50 Uppsala
Sweden

Intended use

Mercodia Animal Insulin Control (Low, Medium, High) is designed to be used as a three-level control for Mercodia Porcine, Ovine, Canine, Equine and two-level control for Bovine Insulin ELISA kits. It is manufactured from human serum and insulin.

For Mercodia assays the control levels are suitable in each method as stated in the table below.

Method	Control		
	Low	Medium	High
Mercodia Porcine Insulin ELISA 10-1200-01	X	X	X
Mercodia Bovine Insulin ELISA 10-1201-01	X	X	-
Mercodia Ovine Insulin ELISA 10-1202-01	X	X	X
Mercodia Canine Insulin ELISA 10-1203-01	X	X	X
Mercodia Equine Insulin ELISA 10-1205-01	X	X	X

Reagents

Each Mercodia Animal Insulin Control Low, Medium, High contains 3 vials of lyophilized serum, each containing 0.5 mL after reconstitution. Values are found below.

Preparation of reagents

Reconstitute each Mercodia Animal Insulin Control (Low, Medium, High) by adding 500 μ L redistilled water. Replace the rubber stopper and wait 5 minutes before vortexing to ensure that the content is thoroughly dissolved.

Shelf life and storage

The expiry date is stated below. Store the complete package at 2-8°C. Reconstituted Controls are stable for 2 weeks at 2-8°C or three months at -20°C. Avoid repeated freezing and thawing.

Procedure

Perform the Mercodia Insulin ELISA assays (Porcine, Bovine, Ovine, Canine and Equine) according to the "Directions for Use", handling Mercodia Animal Insulin Control Low, Medium, High as samples.

Warnings and precautions

Warning! This kit contains reagents that may be infectious!

This kit contains reagents manufactured from human blood components. The source of material have been tested by immunoassay for hepatitis B surface antigen, antibodies for hepatitis C virus and for antibodies for HIV virus and found to be negative. Nevertheless, all recommended precautions for the handling of blood derivatives should be observed. Please refer to HHS Publication No. (CDC) 88-8395 or corresponding local/national guidelines on laboratory safety procedures.

- The Control Low, Control Medium and Control High contain <0.06% 5-chloro-2-methyl-4-isothiazolin-3-one and 2-methyl-2H-isothiazol-3-one (3:1).

The Controls are labeled:



Warning

H317 – May cause an allergic skin reaction.

P280 – Wear protective gloves.

P261 – Avoid breathing vapour.

P272 – Contaminated work clothing should not be allowed out of the workplace.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P333 + P313 – If skin irritation or rash occurs: Get medical attention.

P501 – Dispose of contents and container in accordance with all local, regional, national and international regulations.

Table of values

The concentrations of the control values were determined in 4 replicates per run , on a minimum of 10 occasions, performed by atleast 2 technicians.

The range is determined as: target value ± 3 SD.

Control Low	Lot No.: 33915	Exp.date: 2025-06-30			
			Target value	Range	Units
Mercodia Porcine Insulin ELISA 10-1200-01			5.30	4.51 - 6.10	mU/L
Mercodia Bovine Insulin ELISA 10-1201-01			0.435	0.344 - 0.526	$\mu\text{g/L}$
Mercodia Ovine Insulin ELISA 10-1202-01			0.301	0.238 - 0.364	$\mu\text{g/L}$
Mercodia Canine Insulin ELISA 10-1203-01			5.30	4.51 - 6.10	mU/L
Mercodia Equine Insulin ELISA 10-1205-01			5.30	4.51 - 6.10	mU/L

Control Medium		Lot No.: 33916	Exp.date: 2025-06-30	
		Target value	Range	Units
Mercodia Porcine Insulin ELISA 10-1200-01		17.1	14.5 - 19.7	mU/L
Mercodia Bovine Insulin ELISA 10-1201-01		1.40	1.11 - 1.69	µg/L
Mercodia Ovine Insulin ELISA 10-1202-01		0.830	0.706 - 0.955	µg/L
Mercodia Canine Insulin ELISA 10-1203-01		17.1	14.5 - 19.7	mU/L
Mercodia Equine Insulin ELISA 10-1205-01		17.1	14.5 - 19.7	mU/L

Control High		Lot No.: 33917	Exp.date: 2025-06-30	
		Target value	Range	Units
Mercodia Porcine Insulin ELISA 10-1200-01		61.3	52.1 - 70.5	mU/L
Mercodia Ovine Insulin ELISA 10-1202-01		2.15	1.83 - 2.47	µg/L
Mercodia Canine Insulin ELISA 10-1203-01		61.3	52.1 - 70.5	mU/L
Mercodia Equine Insulin ELISA 10-1205-01		61.3	52.1 - 70.5	mU/L

Determinations performed at Mercodia AB.

The values given are only applicable to the lot number and the expiry date mentioned in the table and to assays manufactured by Mercodia AB. If used with other methods, different values may be obtained.




Good practice dictates that each laboratory establishes its own expected range of values.

Warranty

The performance data presented here was obtained using the procedure indicated. Any change or modification in the procedure not recommended by Mercodia AB may affect the results, in which event Mercodia AB disclaims all warranties expressed, implied or statutory, including the implied warranty of merchantability and fitness for use.

Mercodia AB and its authorized distributors, in such event, shall not be liable for damages indirect or consequential.

Explanation of symbols used on labels

	Expiry date
	Store between 2-8°C
	Lot No.

For technical support please contact: support@mercodia.com